

Citation for published version:

Md Alias, MF, Needham, A, Martinez Hernandez, U, Dehghani-Sanij, AA, Vaidyanathan, R & Weston, C 2018, Walking Activity Recognition with sEMG Sensor Array on Thigh Circumference using Convolutional Neural Network. in *International Conference on Intelligent Systems Engineering (ICISE)*. IEEE.

Publication date:
2018

Document Version
Peer reviewed version

[Link to publication](#)

© 2018 IEEE. This is an author produced version of a paper accepted for publication in 2018 International Conference on Intelligent Systems Engineering (ICISE). Personal use of this material is permitted. Permission from IEEE must be obtained for all other uses, in any current or future media, including reprinting/republishing this material for advertising or promotional purposes, creating new collective works, for resale or redistribution to servers or lists, or reuse of any copyrighted component of this work in other works. Uploaded in accordance with the publisher's self-archiving policy.

University of Bath

Alternative formats

If you require this document in an alternative format, please contact:
openaccess@bath.ac.uk

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Walking Activity Recognition with sEMG Sensor Array on Thigh Circumference using Convolutional Neural Network

Mohd Farid Md Alias*, Ashwin Needham†, Abbas A. Dehghani-Sanij*,

Ravi Vaidyanathan†, Uriel Martinez-Hernandez*, Charles Weston*

**Institute of Design, Robotics and Optimisation, University of Leeds, UK*

{mnmfma, a.a.dehghani-sanij, u.martinez, mn14cw}@leeds.ac.uk

†Biomechatronics Laboratory, Imperial College London, UK

{a.needham15, r.vaidyanathan}@imperial.ac.uk

Abstract—In recognition of walking gait modes using surface electromyography (sEMG), the use of sEMG sensor array can provide sensor redundancy and less rigorous identification of sEMG electrode placements as compared to the conventional sEMG electrode placements right in the middle of muscle bellies. However, the potentially lesser discriminative and noisier sEMG signals from the sEMG sensor array pose the challenge in developing accurate and robust machine learning classifier for walking activity recognition. In this paper, we explore the use of convolution neural network (CNN) classifier with frequency gradient feature derived from EMG signal spectrogram for detecting different walking activities using an sEMG sensor array on thigh circumference. EMG dataset from five healthy subjects and an amputee for five walking activities namely walking at slow, normal and fast speed, ramp ascending and ramp descending are used to train and test the CNN-based classifier. Our preliminary findings suggest that frequency gradient feature can improve the CNN-based classifier performance for walking activity recognition using EMG sensor array on thigh circumference.

Index Terms—electromyography, sensor array, walking gait intent recognition, convolutional neural networks, frequency gradient feature

I. INTRODUCTION

Myoelectric control is a control approach which utilises electromyography (EMG) signals as its control input to recognise muscle activities and subsequently derive the actuator output commands. Myoelectric control has been widely applied in many assistive robotics devices such as exoskeletons [1] and prosthesis [2].

Surface electromyography (sEMG) is a commonly preferred way of acquiring the EMG signals due to its non-invasive nature and ease of use [3]. The sEMG electrodes are typically attached onto the skin areas of the middle of the subject muscle bellies where highest motor unit activations can be detected. Such approach however requires proper palpation procedure to locate the exact locations of those muscles [4]. This can be time consuming and at some point impractical for everyday wearing of any lower-limb assistive robotic devices. To alleviate this problem, alternatively the EMG sensors can be arranged in the form of a sensor array and then be worn

based on certain body parts made as landmarks. For example, an array of EMG sensor can be placed along the circumference of a subject's thigh at a certain distance from the patella. This setting however would need to cope with the crosstalks between muscles and even potentially very low signal due to electrodes being placed on the skin areas for iliotibial tracts or tendons. Hence, pattern recognition strategies used to recognise human activities associated with the EMG signals from these areas would need to be robust to such challenge.

One of the increasingly popular artificial neural network architectures for pattern recognition problem is the convolutional neural network (CNN) [5]. CNN has been successfully applied in the field of image recognition [6], [7]. In such application, the red-green-blue (RGB) image data have been typically used as the input to the network models. To extend the use of CNN for sequential data such as the EMG signals, the CNN needs to be able to detect the temporal patterns within its input data structure.

To fit the spatio-temporal information of the EMG signals analogous to the standard CNN image input representation, the EMG spectrogram can be used as the image input to the CNN in which the time lapse and frequency axes of the graph represent the horizontal and vertical axes of the image input respectively. With such setting, the convolution filters would be able to scan for any temporal patterns as they strides along the horizontal axis of the input matrix.

Here we suggest that the temporal information can be also be potentially augmented by the use of frequency gradient along the frequency axis of the EMG spectrogram. We hypothesise that the frequency gradient can have more discriminative and consistent characteristics over time compared to frequency amplitude. Hence the use of frequency gradient can potentially improve a CNN-based classifier to become more robust toward EMG noise and produce more consistent output over time. In this paper, we propose the use of a classifier based on convolution neural network (CNN) with frequency gradient feature from the EMG spectrogram to recognise different human walking activities based on sEMG measured along the

thigh circumference using a sEMG sensor array.

Our contributions are two, first is to introduce the procedure of applying an array of sEMG sensors along the subject's thigh circumference to acquire sEMG signals from the thigh for activity recognition and potentially to control assistive robotic devices. This procedure involves the identification of electrode locations based on patella as the landmark, which can alleviate the need of rigorous muscle palpation procedure and help to establish more consistent electrode locations for repeated wearing of EMG sensors. Our second contribution is the development of a CNN-based classifier which utilises the frequency gradient feature derived from the EMG spectrograms as the image inputs to the CNN. We compare its performance to another CNN-based classifier which uses EMG spectrogram only as the image inputs and a standard support vector machines-based (SVM-based) classifier as the baseline.

This paper is divided into six sections. Section II briefly explains related works, Section III explains about our EMG data acquisition and analysis procedures, Section IV describes our classification approach, Section V presents our results and discussions respectively, and finally Section VI provides the conclusions of this paper.

II. RELATED WORKS

Biosignal-based sensor arrays have been widely applied in the recognition of human upper limb movements. In [8], three sets of 128-channel EMG sensor arrays placed on the triceps, biceps and forearm areas were used to obtain high density EMG signals for forearm movement recognition. [9] used MMG sensor array on the wrist for identification of finger movements. A hybrid EMG-MMG sensor array was utilised to characterise hand muscle activities during object manipulation tasks.

CNN has been used for classification of hand movements using standard muscle locations as in [10]. [11] use Myo Armband by by Thalmic Inc [12] to obtain the EMG signals at 200 Hz sampling frequency. They then use the EMG spectrogram as the input to their CNN-based classifier. In [13], one-dimensional convolution operation is applied along the frequency band only of the EMG spectrogram within each single time-step.

Applications of sensor array for human activity recognition based on lower-limb movements has been less explored as compared to upper-limb movements. In [14], input signals from an array of eight mechanomyography (MMG) sensors and two inertial measurement units (IMUs) are used to develop classifiers to identify four different walking activities namely walking, stair ascending and stair descending. Whereas [15] uses a neoprene cuff consisting of four piezoresistive force sensors and silicone balloon actuators as a haptic feedback system based on the ground reaction forces. To the best of our knowledge, our work is the first to use the frequency gradient feature derived from the EMG spectrograms as the inputs to the CNN classifier for walking activity recognition using EMG sensor array.




Fig. 1. Electrode placements along the thigh circumference




Fig. 2. Electrode IDs on single thigh

III. EMG DATA COLLECTION

A. Experiment Setup

To evaluate the feasibility of EMG sensor array for walking gait recognition based on thigh muscle activities, we conducted EMG data collection sessions for five different walking activities namely level ground walking at slow, normal and fast speeds, ramp ascending and ramp descending. The data were obtained from ten male subjects of which nine were healthy individuals and one was a transfemoral amputee. The subjects' ages ranged between 19 and 77 years old with average heights of 177.5 ± 7.3 cm and average weights of 74.8 ± 10.6 kg. This study was approved by the University of Leeds Research Ethics Committee and all the subjects had signed their written informed consents before participating in this experiment.

For our EMG sensor array, we used eight Biometrics SCX-1000 EMG sensors. Each sensor can be attached to the skin area with a transparent medical grade tape. It also has a built-in gain of 1000 with bi-polar electrode distance of 15 cm.

The EMG sensors were connected to the National Instruments NI-USB 6212 data acquisition (DAQ) device which was connected to a Windows 10 laptop with a 5 m USB cable. The EMG data stream was recorded using a LabVIEW program in the laptop with a sampling frequency of 1 kHz per channel. Two mechanical switches were also connected to the DAQ hardware to be used by the experiment investigator. One was a latching used to control the remote start and stop of each data logging session and another was a non-latching one to indicate the start and stop of each trial or set.

During each data collection session with a subject, we first identified the circumference location to place the EMG sensor array which was 10 cm from the farthest surface of the subject's patella during sitting position. We then shaved and cleaned skin area along the circumference to improve the EMG signal detection. After that eight Biometrics SCX-1000 EMG sensors were attached along the thigh circumference using medical grade tape at 45° from each other with respect to the centre of the thigh circumference cross-section as shown in Figure 1. The EMG sensors were assigned identification labels from 0 to 7 as in Figure 2. Sensor 0 basically represented biceps femoris, sensor 1 (semitendinosus), 2 (semimembranosus), 3 (gracilis), 4 (vastus medialis), 5 (rectus femoris), 6 (vastus lateralis) and 7 (illotibial tract).

Another main equipment in this experiment was a 5-meter long ramp with inclination of 5° from the ground. The ramp was also installed with aluminium parallel bars as a safety feature for subjects to hold onto in case of becoming imbalanced. For the ramp walking activities, the walking distance was set at 2.5 meters from the lower edge of the ramp. Meanwhile the level ground walking distance was set at 3 meters. Both distances would allow the subjects to achieve at least two full walking strides based on men's average stride length of 88.5 cm [16].

B. Experiment Protocol

Each subject was first instructed to perform two preliminary activities to obtain his maximum voluntary contraction (MVC) values. The first activity was the isometric knee extension in seating position by having the subject's ankle tied to the chair leg with a strap. In the second activity, the subject tried to perform knee flexion in standing position while his ankle was tied to a stationery vertical rod in front of him. The highest EMG value from both activities was then taken as the MVC for each channel. The subject would perform each activity in three sets with two minutes of rest between sets. For each set, the subject would first relax at the starting position for ten seconds to reduce baseline offset followed by performing the movements for six repetitions with ten seconds of rest between each repetition.

The subject then performed one set of level ground walking activity at slow speed to the other end of the walking distance. The next set was performed by returning to the start position. This procedure was repeated for six trials and for another two speeds namely the normal and fast speeds. Two and five minutes of rest were allowed between each trial and activity




Fig. 3. Sample of EMG signals from the sensor array for two consecutive walking trials




Fig. 4. Feature transformation from EMG spectrogram to frequency gradient

respectively. Finally the subject was instructed to walk up and down the ramp for six trials with two minutes of rest between the activities.

C. Data Preprocessing

We first filtered our raw EMG signals using high pass filter at 20 Hz, bandpass filtered them with a fourth order Butterworth filter with the cut-off frequencies of 20 and 450 Hz and then applied notch filter between 59 to 61 Hz to minimise the powerline noise. A sample set of EMG envelopes from the eight EMG electrodes of a subject for two consecutive walking trials are as shown in Figure 3.

D. Feature Extraction

For the CNN-based classifier, we used frequency gradient feature derived from time-frequency graph or spectrogram feature respectively. While these feature selection seems to be very minimal as compared to other EMG-related works [2],




Fig. 5. CNN-based classification approach for walking gait recognition

such a minimal feature extraction approach was intended to test our classifier training with minimal computational load for future real-time control of assistive robotics device. We chose 50 ms and 200 ms respectively as our time window overlap size and main time window size for the frequency gradient and spectrogram features based on earlier works [11], [18], [19] and [20].

The EMG spectrogram of each sensor for a single time step was obtained using short-time Fourier transform (STFT) based on Matlab implementation. To derive the frequency gradient feature from the EMG spectrogram, only the real part of the spectrogram was involved for the computation to minimise computation complexity. Four gradient computation methods have been considered, namely Sobel operator, Prewitt operator, central difference and intermediate difference methods. For our main classifier training of each subject, we have selected Sobel operator based on [17] due to its edge enhancement effects which may enhance the discriminative characteristic of frequency gradient image. During the later part of our research, other gradient operators have also been applied on a single subject's data to compare their computation times. Similar data was used to measure the computation time for STFT and continuous wavelet transform (CWT) to obtain the EMG spectrogram.

Sobel gradient operator along the vertical direction, T_{Sy} is defined as follows [17]:

$$T_{Sy} = \begin{bmatrix} -1 & -2 & -1 \\ 0 & 0 & 0 \\ +1 & +2 & +1 \end{bmatrix} \quad (1)$$

The vertical gradient G_y can be computed from an image input I using Sobel operator using the following formula:

$$G_y = T_{Sy} * I \quad (2)$$

where $*$ refers to convolution operation.

Figure 4 illustrates transformation effect from EMG spectrogram to frequency gradient using Sobel operator along the frequency band axis. Through visual observation, the transformation managed to uncover latent temporal pattern of the EMG spectrogram. The EMG spectrogram image displays more noticeable colour gradient along the frequency band but

colour changes along the time axis would be more useful for feature separability.

IV. CLASSIFICATION APPROACH

A. Convolution Neural Network Architecture

Figure 5 shows our classification approach using a simple CNN architecture. The CNN consists of an eight-dimensional input layer for the eight EMG inputs, a convolutional layer, a rectified linear unit (ReLU) activation layer, a fully-connected layer, a softmax activation layer and the output layer. Each input dimension corresponds to a 129-by-8 matrix of EMG spectrogram or frequency input. The kernel size and the stride size of the convolution filter is 4-by-4 and 4-by-1 respectively. The CNN hyperparameters are set as follows; batch size as 20, number of epoch at 100 and learning rate at 0.005.

B. Classifier Evaluation

The data analysis was conducted in Matlab R2017a on a 64-bit Windows PC with Intel Xeon 3.2 GHz processor and 8 GB RAM. To evaluate our classifier performance, we stratified our data into three sets, namely training set (80%), validation set (5%) and test set (15%). We then reran the classifier training for each subject to obtain the subject-specific classification accuracy. Finally we calculated the average classification accuracy by adding the classification accuracies for all subjects and divide the sum with the number of subjects.

V. RESULTS AND DISCUSSIONS

A. Classification Accuracy

Table I shows the classification results between different classifiers namely CNN-based with frequency gradient features, CNN with spectrogram feature and SVM-based classifier at 200 ms time window. CNN-based classifier with feature gradient feature performs better than CNN-based classifier with spectrogram feature and SVM-based classifier. The use of frequency gradient as the classifier feature helps to increase the classification accuracy by about 6.39% from spectrogram feature. This findings support our assumption that the temporal information in EMG frequency gradient feature can be more discriminative and consistent as compared to EMG spectrogram feature. However, with bigger time window size of the input frame, the frequency gradient advantage reduces and the

feature contributes no better than the EMG spectrogram toward the performance of CNN-based classifier as shown in Figure ???. This can be explained by higher abundance of temporal information in bigger time windows which can be used to derive more feature map by the convolution filter. Yet smaller time window size is more attractive for myoelectric control application to produce faster responses. For the purpose of offline training, the use of longer time windows is useful for better classification performance since more temporal patterns can be detected by the classifier. But question remains as whether the longer time windows can be practical for real-time myoelectric control and further investigation is needed on this issue. Samples of accuracy and loss over the training iteration for one subject and confusion matrix can be found in Figure 6 and Figure ??.

The classification accuracy results however are slightly lower generally achieved accuracy of 80% above in other EMG classification research works [21]. We attribute this to the resultant EMG signal quality due to hardware limitation. Specifically, the EMG electrode tapes that we tested could not continuously attach the EMG electrodes firmly to some subjects' skin areas due to changes in the shapes of the lower third of their thighs during walking movements. Furthermore, the thigh volume changes also has caused some skin tensions on the electrodes and hence movement artifacts to the EMG signals. Better EMG measurements could potentially be achieved with firm placements of the EMG sensor array at different thigh heights.

By measuring the EMG signals on implicit muscle locations as we performed, the acquired EMG signals might also be more prone to crosstalks between muscles. However, even with the existing noisy measurements, the results suggest that the inputs from EMG sensor array could help the classifier to leverage those effects to a certain extent rather than behaving as a random classifier. In addition, with more data samples, the CNN-based classifier can be further trained and tested for improved classification accuracy.

B. Computational Costs

Table II shows the computational times of different image gradient methods, spectrogram and CWT for single input sample based on Matlab implementation. Spectrogram computation using either STFT or CWT was found to be substantially longer than frequency gradient computation. Computation time for the Sobel method is comparable to the Prewitt method but both are slightly slower than central and intermediate difference methods. The differences can be considered trivial in offline training. However, if computational efficiency remains crucial, then the latter two methods can be considered for gradient computation.

C. Limitations and Future Works

Our current CNN-based classifier development here is mainly intended as a proof of concept particularly with regards to the use of sEMG sensor array on thigh and frequency gradient feature. Hence the classifier only consists of a simple

TABLE I
COMPARISON OF CLASSIFICATION ACCURACIES AT 200 MS TIME WINDOW

Classifier	Accuracy
CNN with frequency gradient feature	70.93 %
CNN with spectrogram feature	64.54 %
SVM	67.8 %




Fig. 6. Accuracy and loss over the training iteration for a subject

CNN architecture without additional optimisation steps such as dropout, batch normalisation or momentum. More complex CNN architecture may be developed in the future.

The final goal of this research is to develop a novel ANN-based myoelectric control scheme of an assistive robotic device. The developed classifier will be used as part of the high-level control scheme that can be connected to the middle and low-level controllers. For example, higher level knowledge of current walking activity type would be used to predict the suitable torque responses for such walking condition.

Our future works will also include improvement of the classifier for subject-independent setting and experimentation with different EMG sensor array locations namely the higher parts of the thigh. Higher number of subjects and evaluation of classifier performance over time will also be considered in future experiments.

VI. CONCLUSIONS

As the conclusion, sEMG sensor array application on thigh circumference can be potentially used as an alternative approach to standard electrode placement procedure. However, the inherent crosstalks and weaker signals on certain locations need to be addressed with the use of more robust classifier and discriminative features. Frequency gradients of EMG spectrogram is one promising feature to serve as the image input to a CNN-based classifier.

TABLE II
AVERAGE COMPUTATION TIMES OF FREQUENCY GRADIENT AND
SPECTROGRAM FEATURES FOR A SINGLE ONE-DIMENSIONAL INPUT
SAMPLE

Computation methods	Computation time (s)
Spectrogram:	
STFT	0.1782
CWT	0.1907
Frequency Gradient:	
Sobel method	0.0007
Prewitt method	0.0008
Central difference method	0.0003
Intermediate difference method	0.004

ACKNOWLEDGEMENT

The first author would like to thank International Islamic University Malaysia (IIUM) for sponsoring his PhD fellowship for this research. The team also wish to thank the following parties; Mechatech Ltd for sponsoring the research, Affective Engineering Lab of the University of Leeds for becoming the venue for the experiment and Biomechanics Lab of Imperial College London for the collaboration work.

REFERENCES

- [1] H. Kawamoto, S. Lee, S. Kanbe, and Y. Sankai, "Power assist method for hal-3 using emg-based feedback controller," in *Systems, Man and Cybernetics, 2003. IEEE International Conference on*, vol. 2, pp. 1648–1653, IEEE, 2003.
- [2] M. A. Oskoei and H. Hu, "Myoelectric control systemsa survey," *Biomedical Signal Processing and Control*, vol. 2, no. 4, pp. 275–294, 2007.
- [3] P. Konrad, "The abc of emg," *A practical introduction to kinesiological electromyography*, vol. 1, pp. 30–35, 2005.
- [4] H. J. Hermens, B. Freriks, R. Merletti, D. Stegeman, J. Blok, G. Rau, C. Disselhorst-Klug, and G. Hägg, "European recommendations for surface electromyography," *Roessingh research and development*, vol. 8, no. 2, pp. 13–54, 1999.
- [5] A. Krizhevsky, I. Sutskever, and G. E. Hinton, "Imagenet classification with deep convolutional neural networks," in *Advances in neural information processing systems*, pp. 1097–1105, 2012.
- [6] K. Simonyan and A. Zisserman, "Very deep convolutional networks for large-scale image recognition," *CoRR*, vol. abs/1409.1556, 2014.
- [7] K. He, X. Zhang, S. Ren, and J. Sun, "Deep residual learning for image recognition," *CoRR*, vol. abs/1512.03385, 2015.
- [8] M. Jordani, M. Rojas-Martinez, M. A. Maanas, J. F. Alonso, and H. R. Marateb, "A novel spatial feature for the identification of motor tasks using high-density electromyography," *Sensors*, vol. 17, no. 7, p. 1597, 2017.
- [9] R. Booth and P. Goldsmith, "Validation of a piezoelectric sensor array for a wrist-worn muscle-computer interface," *CMBES Proceedings*, vol. 39, no. 1, 2016.
- [10] M. Atzori, M. Cognolato, and H. Müller, "Deep learning with convolutional neural networks applied to electromyography data: a resource for the classification of movements for prosthetic hands," *Frontiers in neurorobotics*, vol. 10, p. 9, 2016.
- [11] U. Côté-Allard, C. L. Fall, A. Campeau-Lecours, C. Gosselin, F. Lavolette, and B. Gosselin, "Transfer learning for semg hand gestures recognition using convolutional neural networks,"
- [12] Thalmic Labs Inc., "Myo Gesture Control Armband — Wearable Technology by Thalmic Labs," 2016. (Accessed on 01/31/2018).
- [13] P. Xia, J. Hu, and Y. Peng, "Emg-based estimation of limb movement using deep learning with recurrent convolutional neural networks," *Artificial organs*, 2017.
- [14] A. Needham and Ravi, "Mechanomyography..," *A practical introduction to kinesiological electromyography*, vol. 1, pp. 30–35, 2017.
- [15] R. E. Fan, M. O. Culjat, C. H. King, M. L. Franco, R. Boryk, J. W. Bisley, E. Dutson, and W. S. Grundfest, "A haptic feedback system for lower-limb prostheses," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 16, pp. 270–277, June 2008.
- [16] P. E. Martin and R. C. Nelson, "The effect of carried loads on the walking patterns of men and women," *Ergonomics*, vol. 29, no. 10, pp. 1191–1202, 1986.
- [17] W. Gao, X. Zhang, L. Yang, and H. Liu, "An improved sobel edge detection," in *Computer Science and Information Technology (ICCSIT), 2010 3rd IEEE International Conference on*, vol. 5, pp. 67–71, IEEE, 2010.
- [18] M. Zardoshti-Kermani, B. C. Wheeler, K. Badie, and R. M. Hashemi, "Emg feature evaluation for movement control of upper extremity prostheses," *IEEE Transactions on Rehabilitation Engineering*, vol. 3, no. 4, pp. 324–333, 1995.
- [19] L. H. Smith, L. J. Hargrove, B. A. Lock, and T. A. Kuiken, "Determining the optimal window length for pattern recognition-based myoelectric control: balancing the competing effects of classification error and controller delay," *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, vol. 19, no. 2, pp. 186–192, 2011.
- [20] B. Hudgins, P. Parker, and R. N. Scott, "A new strategy for multifunction myoelectric control," *IEEE Transactions on Biomedical Engineering*, vol. 40, no. 1, pp. 82–94, 1993.
- [21] J. Yousefi and A. Hamilton-Wright, "Characterizing emg data using machine-learning tools," *Computers in biology and medicine*, vol. 51, pp. 1–13, 2014.